

## Installation Guide


### Focus Flush Mount Kit for Spacelabs (90499) 2-Bay Module Housing

*The Purpose of this guide is to describe mounting the Focus Flush Mount Kit:*

#### Parts Reference

The following parts and hardware are included in this installation kit (hardware not shown):

Item #	Description	Qty.
1	Flush Mount	1
2	90499 Adapter Plate	1
3	Utility Hook	1
4	8-32 x 5/8" Pan Head Machine Screw (PHMS)	4
5	#8 Flat Washer	4
6	8-32 x 1/4" Flat Head Machine Screw (FHMS)	4
7	8-32 x 3/16" PHMS	2
8	1/8" Hex Wrench	1




#### Tools Required:

- 1/8" Hex Wrench (provided)
- Philips screwdriver (not provided)


#### Installing Flush Mount

Guide the Flush Mount Slide into the top of the Channel (below left). Move the Flush Mount to desired mounting position and tighten the (2) set screws using an 1/8" hex wrench (below right).




## Installing 90499 Module Housing

1. Align the (4) mounting holes on the adapter plate with the (4) threaded mounting holes on the bottom of the module housing. Using a Phillips screwdriver, secure mounting adapter to module housing with (4) #8 flat washers and (4) 8-32 x 5/8" PHMS. Rotate the module housing and slide it into the flush mount.




2. Align the (4) mounting holes on the Flush Mount with the (4) threaded mounting holes on the adapter plate. Using a Phillips screwdriver, secure the module housing with (4) 8-32 x 1/4" FHMS as shown below.



## Utility Hook

Using a Phillips screwdriver, attach the utility hook to the side of the flush mount with (2) 8-32 x 3/16" PHMS.



## Routine Maintenance

Periodically inspect all fasteners associated with the mounting assembly. Tighten or Adjust as necessary for optimal operation and safety.

### Cleaning the Mounting Assembly

The mounting assembly may be cleaned with most mild, non-abrasive solutions commonly used in the hospital environment (e.g. diluted bleach, ammonia, or alcohol solutions).

The surface finish will be permanently damaged by strong chemicals and solvents such as acetone or trichloroethylene.

Steel wool or other abrasive material should never be used.

Damaged caused by the use of unapproved substances or processes will not be warranted. We recommend testing of any cleaning solution on a small area of the mounting assembly that is not visible to verify compatibility.

Never submerge or allow liquids to enter the mounting assembly. Wipe any cleaning agents off the mounting assembly immediately using a water-dampened cloth. Dry the assembly thoroughly after cleaning.

 **CAUTION:** GCX makes no claims regarding the efficacy of the listed chemicals or processes as a means for controlling infection. Consult your hospital's infection control officer or epidemiologist. To clean or sterilize mounted devices or accessory equipment, refer to the specific instructions delivered with those products.